



November 11, 2008

ASW Engineering
ATTN: Victor Sanchez
RE: EZ View Gauge

To Whom It May Concern:

In regards to Hedland EZ-View Test Kit, part number H623-010, specifically the needle gauge, we are unable to provide a specification sheet from the original manufacturer. We obtain these gauges from a supplier, who in turn, buys them from various manufacturers. The specifications we require on these gauges is a maximum reading pressure of 160 PSI, with $\pm 3\%$ accuracy in both the lower and upper $\frac{1}{4}$ portions of the scale, and $\pm 2\%$ accuracy in the middle portion of the scale. Unfortunately, we have no way of knowing definitively from which company each specific gauge came.

If you have any other questions, please contact me.

Sincerely,

Jen Wehrli-Mead



Racine Flow Meter Group

8635 Washington Avenue, Racine, WI 53406-3738 U.S.A. • hedlandsales@racinefed.com
Tel: 262-639-6770 • 800-433-5263 • Fax: 262-639-2267 • 800-245-3569 • www.hedland.com